

# Service Repair Sheet SRS68824

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Lauren Ratcliffe

Alder Hey Childrens Hospital

Clinical Engineering

Administrator

0151 2525293

01512284811

00012595

lauren.ratcliffe@alderhey.nhs.uk

26/Jun/2024

Robert Connor

Viamed

Warranty

## VIAMEDclean

### Goods In Only

Decontamination  
certificate provided  
by customer

Cleaned by Viamed,  
if no declaration certificate  
from customer

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

### Goods Out Only

Cleaned by Viamed before  
returning to customer

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

Notes 21/Jun/2024 Steve Hardaker

21/Jun/2024 Steve Hardaker

Returning a pair of R-43V, s/n V110449/V110491. The contact ring fell out of V110449 before installation.

Replace ASAP under warranty.

26/Jun/2024 Robert Connor

Received 2 x R43V s/n V110491, V110449.

Ready For quote

Repair Complete Signed

CCN 28.6.24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36509	Oxygen Sensor	0110043	V110449/V110491	Yes

0140013 x 1 FOC  
old S/N, SRS, SRN

0110043 x 1 FOC  
New S/N, SRS, SRN

UPS x 1 FOC.