

Service Repair Sheet 68812

Contact Name	Madisa Adnani
Company/ Hospital Name	Addenbrookes Hospital
Direct Phone	01223245151
General Phone	01223245151
Opera Account	00000860
Email	madisa.adnani1@nhs.net
Order Number	P1000265889
Date Received	24/Jun/2024
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN36507	1410000	PR01576A11	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
