

Service Repair Sheet 68831

Contact Name

Nathaniel Ford

Company/ Hospital Name

Isle of Wight NHS Primary Care

Department

Medical Electronics Dept

Position

Medical Devices Coordinator & Senior Technician

Direct Phone

01983822099

Opera Account

00004205

Email

nathanielford@nhs.net

Order Number

282167922

Date Received

04/Jul/2024

Booked in By

Robert Connor

Repair	Ref	S/N	Equipment Type
SRN36519	1410000	PR03072A10	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.