



VIAMED

Cathy Green

Viamed Limited · 15 Station Road · Cross Hills
Keighley · West Yorkshire · BD20 7DT · United Kingdom
Tel: +44 (0)1535 634 542 · Fax: +44 (0)1535 635 582
Email: info@viamed.co.uk Website: www.viamed.co.uk

Service Repair Sheet 68854

Contact Name

Company/ Hospital Name

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Ramesh Shanmugam

Pilgrim Hospital

01205445205

01205364801

00000600

Ramesh.Babu@ulh.nhs.uk

U033474

22/Jul/2024

Robert Connor

Repair	Ref	S/N	Equipment Type
SRN36536	0310030	VMDHO8	Tom Thumb

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1	0380000
Qty: 1	0330208

Tom Thumb TT480 VMDHO8 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

I added a 10mm inlet as it was missing.