Service Repair Sheet SRS68830 Vincent Wu Contact Name Hui-Li Medical Equipment Limited VIAMEDclean Company/ Hospital Name Goods In Only Department Decontamination certificate provided Position 0088 634 910790 by customer Direct Phone Cleaned by Viamed, if no declaration certificate from customer +886 34910790 General Phone 00006036 Signed: -9/7/24 Opera Account vince.wu@huilimedical.com Email Order Number **Goods Out Only** 09/Jul/2024 Cleaned by Viamed before Date Received Robert Connor returning to customer Booked in By Signed: Viamed Main Company Quote Type Return Notes 28/Jun/2024 Ryan Swaine 28/Jun/2024 Ryan Swaine VM-2500-S - likely to need a new ISA module, please quote for repair. 09/Jul/2024 Robert Connor Received 1 x VM-2500-S s/n 70740154, with silicone cover. CGNEEN 26.7.24 Proforma USD Dist Ready For quote Repair Complete Signed

V

RC

USA Plug

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36526	Capnograph	4410520	70740154	2

\$90 Labo. 4440000 x 1/2 hr & Ego per hr SIN, SRS, SRN 4480520 × 1 2 8450 \$ 131.25 SRS, SEN. 4420504 × 1 0 81,049.75 \$1,282.50 SES, SEN 4400070 × 1 2 650.30 \$73.76. ses, sen 4420598 × 1 2 642.95 \$ 43.80 525, 52N 4430610 4430607 × 10 6335 \$ 6.75. SRS, SAN UPS + 1 2 50.