



Michael Lamb <viamed.michael.lamb@gmail.com>

MISSING PARTS FROM DE ANALIZERS

procurement <procurement@cascoantiguo.com>

6 August 2024 at 15:45

To: "michael.lamb@vandagraph.co.uk" <michael.lamb@vandagraph.co.uk>

Cc: "ryan.swaine@vandagraph.co.uk" <ryan.swaine@vandagraph.co.uk>

Hello Michael & Ryan,

We have checked inside the oxygen analyzer because a client noticed it, and there are some missing parts. Attached is the missing part.

We have a total of 6 analyzers that are missing this. Can you please send us 6 of these B-50057 Flow Divertor with O ring.

Best regards



BUCEO | OCEANOGRAFÍA | MILITAR | RESCATE

Javier Vicioso

Procurement Department

procurement@cascoantiguo.com

+(34) 617 80 51 69

C/ De Senda Galiana, C-18, 28821, Coslada,
Madrid, Españawww.cascoantiguo.com

2 attachments**B-50057 Flow
Divertor**
(9711002) includes O ring**Captura (1).PNG**
63K**20240806_163953.jpg**
4349K