

Service Repair Sheet 68852

Contact Name

Ester Felder

Company/ Hospital Name

MTTBV

Direct Phone

0031 5754 52919

General Phone

310575452919

Opera Account

00007497

Email

e.felder@mttnl.com

Date Received

29/Jul/2024

Booked in By

Robert Connor

Repair	Ref	S/N	Equipment Type
SRN36561	1410000	PR0531A11	Foetal Simulator

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
