

Service Repair Sheet 68827

Contact Name

Umer Tariq

Company/ Hospital Name

Umer Tariq

Direct Phone

07535 480529

General Phone

07535480529

Opera Account

CID23684

Email

umertariq616@gmail.com

Order Number

proforma paid for

Date Received

29/Jul/2024

Booked in By

Robert Connor

| Repair | Ref | S/N | Equipment Type |
|----------|---------|------------|------------------|
| SRN36565 | 1410000 | PR03507A14 | Foetal Simulator |

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1 1480000

Qty: 1 1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
