



DECLARATION OF CONFORMITY

Description: **DB200381 Oxygen Sensor**
Part number: **8010050**

Manufacturer: **Vandagraph Sensor Technologies Ltd.**
15 Station Road
Cross Hills
Keighley
West Yorkshire BD20 7DT
United Kingdom

Specification: **OEM Specification Datasheet – 8010050, v2.0 dated 6th May 2019**

Standards: **ISO 9001-2015**

JFD purchase order number: 12403075-4

Serial number/s	Manufacture – Year/Month
106646– 106681 (qty 36)	2024/08

I, the undersigned, hereby declare that the oxygen sensors listed above conform to the stated specifications.

Signature:

Name: **S Nixon**

Position: **Director**

Date: **2nd August 2024**



VERIFIED
CERTIFICATE

CHECKED BY: **C. WALLS** DATE: **14/8/24**
RECEIVER # **306179** OCR #
WO/PO # **P012403 07511**
JFD PART NUMBER **DB200381 qty 36**
TRACEABLE MARK **AS HIGHLY H TEO**

Purchase order

12403075-4

L SnT

7 SnT

Date	Part number	Part number	Serial number	Output	Zero offset voltage	
	VST	JFD		mV	μ V	
04/07/2024	8010050	DB200381	106646	7.8	50	4.13
04/07/2024	8010050	DB200381	106647	7.9	34	4.37
04/07/2024	8010050	DB200381	106648	7.8	55	4.77
04/07/2024	8010050	DB200381	106649	8.0	47	3.34
04/07/2024	8010050	DB200381	106650	7.9	45	6.42
04/07/2024	8010050	DB200381	106651	7.9	25	5.05
04/07/2024	8010050	DB200381	106652	7.8	73	5.08
04/07/2024	8010050	DB200381	106653	7.8	33	3.42
04/07/2024	8010050	DB200381	106654	8.0	42	4.10
04/07/2024	8010050	DB200381	106655	7.7	42	4.07
04/07/2024	8010050	DB200381	106656	7.9	21	4.20
04/07/2024	8010050	DB200381	106657	7.9	31	3.61
04/07/2024	8010050	DB200381	106658	7.8	30	4.73
04/07/2024	8010050	DB200381	106659	7.8	47	4.11
04/07/2024	8010050	DB200381	106660	7.8	42	3.84
04/07/2024	8010050	DB200381	106661	8.0	38	5.20
04/07/2024	8010050	DB200381	106662	7.9	19	4.42
04/07/2024	8010050	DB200381	106663	7.8	40	3.77
04/07/2024	8010050	DB200381	106664	7.9	26	3.19
04/07/2024	8010050	DB200381	106665	7.9	40	4.36
04/07/2024	8010050	DB200381	106666	7.9	45	4.51
04/07/2024	8010050	DB200381	106667	7.8	40	3.88
04/07/2024	8010050	DB200381	106668	7.8	20	4.55
04/07/2024	8010050	DB200381	106669	7.9	26	4.23
04/07/2024	8010050	DB200381	106670	7.7	32	3.41
04/07/2024	8010050	DB200381	106671	7.8	45	5.10
04/07/2024	8010050	DB200381	106672	7.8	39	5.61
04/07/2024	8010050	DB200381	106673	7.8	45	4.25
04/07/2024	8010050	DB200381	106674	7.9	31	5.01
04/07/2024	8010050	DB200381	106675	7.8	30	5.60
04/07/2024	8010050	DB200381	106676	7.8	37	3.72
04/07/2024	8010050	DB200381	106677	7.9	55	5.68
04/07/2024	8010050	DB200381	106678	7.7	40	5.23
04/07/2024	8010050	DB200381	106679	7.8	39	3.77
04/07/2024	8010050	DB200381	106680	7.7	45	4.38
04/07/2024	8010050	DB200381	106681	7.8	41	4.37



Invoice number: RST150116
Shipment date: 2nd August 2024
JFD reference: 12403075-4

**Magnetic signature analysis of production oxygen sensors
through JFD QA goods-in procedures**

36 x 8010050 Oxygen Sensor – JFD DB200381

Once the analysis has been carried out, please complete the following and email back to VST:

info@vandagraphst.com

Result (PASS/FAIL):

Qty 26 PASS 15nt

Qty 10 FAIL 75nt, AS HIGHLIGHT OVERLEAF

5.01nt thru 6.42nt

Name: CAUM WALS

Position: QUALITY ASSURANCE ENGINEER

Signed by:
On behalf of JFD



Date: 14 JUL 2024

Vandagraph Sensor Technologies Limited
15 Station Road, Cross Hills, Keighley, West Yorkshire BD20 7DT, United Kingdom
Tel +44 (0) 1535 634900 Fax +44 (0) 1535 635582 e-mail: info@vandagraphst.com
www.vandagraphst.com

Company registered in England, No. 03116432