



DECLARATION OF CONFORMITY

Description: **DB200381 Oxygen Sensor**
Part number: **8010050**

Manufacturer: **Vandagraph Sensor Technologies Ltd.**
15 Station Road
Cross Hills
Keighley
West Yorkshire BD20 7DT
United Kingdom

Specification: **OEM Specification Datasheet – 8010050, v2.0 dated 6th May 2019**

Standards: **ISO 9001-2015**



VERIFIED
CERTIFICATE

CHECKED BY **DM** DATE **01/08/24**
RECEIVER # **306139** OCR # **136**
WO/PO # **12403075/1**
JFD PART NUMBER **DB200381**
TRACEABLE MARK **SN'S HANDLED**
BAOW

JFD purchase order number: 12403075-2

Serial number/s	Manufacture – Year/Month
106574– 106609 (qty 36)	2024/08

I, the undersigned, hereby declare that the oxygen sensors listed above conform to the stated specifications.

Signature:

Name:

S Nixon

Position:

Director

Date:

1st August 2024

Purchase order

12403075-2

L5nT

75nT

Date	Part number	Part number	Serial number	Output	Zero offset voltage	
	VST	JFD		mV	μ V	
02/07/2024	8010050	DB200381	106574	8.0	37	61.70
02/07/2024	8010050	DB200381	106575	7.8	25	5.81
02/07/2024	8010050	DB200381	106576	7.8	62	4.79
02/07/2024	8010050	DB200381	106577	7.9	12	4.21
02/07/2024	8010050	DB200381	106578	8.1	93	4.03
02/07/2024	8010050	DB200381	106579	7.7	9	4.08
02/07/2024	8010050	DB200381	106580	7.5	56	4.04
02/07/2024	8010050	DB200381	106581	7.9	20	3.84
02/07/2024	8010050	DB200381	106582	7.7	100	3.96
02/07/2024	8010050	DB200381	106583	7.8	62	4.02
02/07/2024	8010050	DB200381	106584	7.8	45	5.02
02/07/2024	8010050	DB200381	106585	7.9	25	4.25
02/07/2024	8010050	DB200381	106586	7.8	107	4.05
02/07/2024	8010050	DB200381	106587	7.7	62	3.51
02/07/2024	8010050	DB200381	106588	7.8	42	4.43
02/07/2024	8010050	DB200381	106589	7.8	31	4.01
02/07/2024	8010050	DB200381	106590	7.8	78	4.54
02/07/2024	8010050	DB200381	106591	8.0	37	3.61
02/07/2024	8010050	DB200381	106592	7.8	43	4.02
02/07/2024	8010050	DB200381	106593	7.9	40	4.74
02/07/2024	8010050	DB200381	106594	7.8	44	5.82
02/07/2024	8010050	DB200381	106595	7.7	29	4.43
02/07/2024	8010050	DB200381	106596	7.8	38	5.38
02/07/2024	8010050	DB200381	106597	7.8	50	5.03
02/07/2024	8010050	DB200381	106598	7.8	43	5.06
02/07/2024	8010050	DB200381	106599	8.0	35	4.22
02/07/2024	8010050	DB200381	106600	7.9	34	4.10
02/07/2024	8010050	DB200381	106601	7.9	40	4.11
02/07/2024	8010050	DB200381	106602	7.9	41	3.71
02/07/2024	8010050	DB200381	106603	7.8	38	4.00
02/07/2024	8010050	DB200381	106604	7.9	29	5.01
02/07/2024	8010050	DB200381	106605	8.0	73	4.27
02/07/2024	8010050	DB200381	106606	7.8	33	4.30
02/07/2024	8010050	DB200381	106607	7.8	63	3.72
02/07/2024	8010050	DB200381	106608	7.8	31	4.74
02/07/2024	8010050	DB200381	106609	7.9	62	3.91



Invoice number: RST150114
Shipment date: 1st August 2024
JFD reference: 12403075-2

**Magnetic signature analysis of production oxygen sensors
through JFD QA goods-in procedures**

36 x 8010050 Oxygen Sensor – JFD DB200381

Once the analysis has been carried out, please complete the following and email back to VST:

info@vandagraphst.com

Result (PASS/FAIL):

QTY 28 PASS 25nt

QTY 8 FAIL 75nt, AS HIGHLIGHTED
OVERLEAF

5.01nt THRU 5.87nt 61.70nt

Name: CALUM WALKER
Position: QUALITY ASSURANCE ENGINEER

Signed by:
On behalf of JFD



Date: 14 JUL 2024

Vandagraph Sensor Technologies Limited
15 Station Road, Cross Hills, Keighley, West Yorkshire BD20 7DT, United Kingdom
Tel +44 (0) 1535 634900 Fax +44 (0) 1535 635582 e-mail: info@vandagraphst.com
www.vandagraphst.com

Company registered in England, No. 03116432