



VIAMED

Viamed Limited · 15 Station Road · Cross Hills
Keighley · West Yorkshire · BD20 7DT · United Kingdom
Tel: +44 (0)1535 634 542 · Fax: +44 (0)1535 635 582
Email: info@viamed.co.uk Website: www.viamed.co.uk

Service Repair Sheet 68868

Contact Name	Arlene Gallagher
Company/ Hospital Name	Forth Valley Royal Hospital
Department	Medical Physics Dept
Position	Ms Gallagher
Direct Phone	01324 566953
General Phone	01324 566000
Opera Account	00004880
Email	arlene.gallagher@nhs.scot
Order Number	VMF14462722
Date Received	05/Aug/2024
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN37035	1410000	PR02724A15	Foetal Simulator	Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.