

Service Repair Sheet 68847

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By

Mohamed Azat
Stepping Hill Hospital
EBME Department
Senior EBME Engineer
01614195447
01614831010
00004891
Mohamed.Azat@stockport.nhs.uk
322000733
22/Jul/2024
Robert Connor

Repair Ref S/N
SRN36537 2510000 H0002128

Equipment Type
Microstim DB3

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1 2540010
Qty: 1 2520000

The Microstim DB3 has been repaired. The patient lead has been replaced.