

Service Repair Sheet 68877

Contact Name

Company/ Hospital Name

Position

Direct Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Muhammad Sami Uddin Naseer

Ergea UK and Ireland Limited

Field Service Engineer (East)

0118 900 8140

00013122

sami.naseer@ergeagroup.com

AS01-PO049054-1

21/Aug/2024

Robert Connor

Repair	Ref	S/N	Equipment Type
SRN37065	2510000	M0004677	Microstim DB3

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1 2540010

Qty: 1 2520000

The Microstim DB3 has been repaired. The patient lead has been replaced.