

Cathy Green

Equipment for Repair

Date:

6/9/24

To:

VIAHWS

Repair Reference - 61433

Vet Direct

Units 9 & 10, Gateway West

Newburn Riverside

Newcastle upon Tyne

NE15 8NZ

Tel: 0191 2642705

Email: repairs@vet-direct.com

The enclosed product has been sent to you for inspection. Please provide estimate to repair - please send estimate to repairs@vet-direct.com

Product: ENVITEC OXIQUANT 5 ANALYSER

Serial No:

A110805/A106106

Accessories:

—

Quantity:

1

Condition:

USED

Work required:

WONT GO BELOW 91
CANT CALIBRATE TO 20.9

Warranty:

Yes / No

Purchase DATE
? 2/5/23

PO Number: 125474

Quotation required?

Yes / No

IF NOT UNDER WARRANTY

Supplier action required:

Repair

Replace / Credit

Vet Direct

...on the pulse