



Service Repair Sheet 68899

Contact Name	Aneta Szubarga
Company/ Hospital Name	Euromed Medical Solution
Direct Phone	0048618439059
General Phone	603243071
Opera Account	00006582
Email	a.szubarga@euromed.net.pl
Order Number	payment recd
Date Received	19/Sep/2024
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN37086	1410000	PR02724A13	Foetal Simulator	Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
