



Kate Griffiths &lt;viamed.kate.griffiths@gmail.com&gt;

**Re: Oxygen sensor order**

1 message

**majdi majdi** <dr\_majdi\_wanis@yahoo.com>  
Reply-To: majdi majdi <dr\_majdi\_wanis@yahoo.com>  
To: kate.griffiths@viamed.co.uk

4 October 2024 at 10:22

Good morning  
Actually i have 52 ventilator manufactured in 2020  
Message was low voltage oxygen sensor  
I think it need to replace?  
Please send me last price invoice 10 pcs as trial

Best regard

Yahoo Mail: Search, Organize, Conquer

On 2024 أكتوبر 4, الجمعة at 11:17 ص, Kate Griffiths  
<kate.griffiths@viamed.co.uk> wrote:

Good morning,

Thank you for your reply.

**We recommend the R-22MEDV as the most directly compatible sensor.**

The sensor S+VOX has an output of 8.0mV to 13.0mV in air.

The Teledyne R-22MED has an output of 7.0mV to 13.0mV in air and will work in most cases, but some R-22MED could fail to calibrate if they are in the 7-8mV range.

The best option is the R-22MEDV, which has an output of 9.0mV to 13.0mV in air, which is within the output range of the S+VOX and therefore directly compatible.

**Would you like a proforma invoice for 10 x R-22MEDV 0110021 at £34 each EXW?**

Kind regards

Kate Griffiths

<http://www.viamed.co.uk>  
Email [kate.griffiths@viamed.co.uk](mailto:kate.griffiths@viamed.co.uk)  
Tel: +44 (0)1535 634542  
Fax: +44 (0)1535 635582

Twitter: [twitter.com/ViamedLtd](https://twitter.com/ViamedLtd)  
Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office.  
Telephone calls to the main office will be answered, but please continue to use email where possible.

E-mail Disclaimer

-----  
GDPR / Privacy Policy available at [www.viamed.co.uk/policy](http://www.viamed.co.uk/policy)  
The information in this email is confidential and may be legally privileged.  
It is intended solely for the addressee. Access to this email by anyone else is unauthorized.  
If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

-----

On Thu, 3 Oct 2024 at 15:47, majdi majdi <[dr\\_majdi\\_wanis@yahoo.com](mailto:dr_majdi_wanis@yahoo.com)> wrote:

Good evening

The email picture are all for aeonmed VG 70 ventilator

Waiting for invoice

Best regards

Yahoo Mail: [Search](#), [Organize](#), [Conquer](#)

On 2024 , 3 أكتوبر , الخميس at 4:17 م, majdi majdi  
<[dr\\_majdi\\_wanis@yahoo.com](mailto:dr_majdi_wanis@yahoo.com)> wrote:

Good afternoon

Plz check the following o2 sensor of different company

Are they same?

Please send me invoice with 10 pcsas with last price

