

Certificate of Calibration

Certificate number: 1456919078

V1000 Foetal Heart Simulator

Part Number: 1410000

Serial number: PR02094A20

This certificate reports recorded values for the instrument specified above.

The output rates were measured in accordance with procedure:
VM3COP55.01, using test equipment CE0185 - calibration traceable to UKAS
calibration standards.

V1000 BPM setting	Period values stated in milliseconds				Result PASS/FAIL
	Target	Lower limit	Upper limit	Actual	
60	1000.00	990.00	1010.00	993.42	PASS
90	666.70	660.03	673.37	665.85	PASS
120	500.00	495.00	505.00	500.19	PASS
150	400.00	396.00	404.00	400.25	PASS
180	333.30	329.97	336.63	332.72	PASS
210	285.70	282.84	288.25	285.92	PASS

Signed: _____

Name: Phil Crossley

Date: 2nd March 2016