



Kate Griffiths <viamed.kate.griffiths@gmail.com>

Re: Payment details

1 message

Sudath Weeraman <sudathweeraman17@gmail.com>

7 October 2024 at 12:07

To: Kate Griffiths <kate.griffiths@viamed.co.uk>

Hi Kate,

Thank you for your email,

I am writing to inform you, we have 02 warranty replacement oxygen sensor model R-42V S/N - V10151 / V101762

These both sensors fault are indicate oxygen percentage is minus valve and calibration failed

Please send a return reference number for above sensors under warranty.

Thank you,

Best regards,

Sudath Weeraman
Prime Lanka Associates
Sri Lanka.

On Mon, Oct 7, 2024 at 4:05 PM Kate Griffiths <kate.griffiths@viamed.co.uk> wrote:

Many thanks Sudath. It is much appreciated.

Kind regards

Kate Griffiths

<http://www.viamed.co.uk>

Email kate.griffiths@viamed.co.uk

Tel: +44 (0)1535 634542

Fax: +44 (0)1535 635582

Twitter: twitter.com/ViamedLtd

Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office. Telephone calls to the main office will be answered, but please continue to use email where possible.