

Service Repair Sheet SRS68923

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Abbey McAnaney

Vet Direct Services Ltd

Technical Support RVN

01912642705

01912642705

00012824

tech@vet-direct.com

07/Oct/2024

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒Cleaned by Viamed,
if no declaration certificate
from customer ☒Signed: ACDate: 7/10/24

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 07/Oct/2024 Robert Connor

07/Oct/2024 Robert Connor

Returned 1 x OxiQuant S s/n A106337, with OOI103 s/n A111126, flue tee, flow diverter and oxygen stem. Customer reports unit will not calibrate, and will not go below 90%.

Ready For quote

Repair Complete Signed

CCNEON 8-10-24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37126	Oxygen Analyser	0111275	A106337	Yes
SRN37127	Oxygen Sensor	0110660	A111126	

} part of
same unit

0140137 x 1 FOC

S/N, SRS, SRN

Includes oxygen sensor.

S/N, SRS, SRN - NO fault found.

UPS x 1 FOC