

Equipment for Repair

Date: 4/10/24
To: VIANCO

Repair Reference - 64922

Vet Direct
Units 9 & 10, Gateway West
Newburn Riverside
Newcastle upon Tyne
NE15 8NZ

Tel: 0191 2642705

Email: repairs@vet-direct.com

The enclosed product has been sent to you for inspection. Please provide estimate to repair - please send estimate to repairs@vet-direct.com

Product:	OXIQUANT 5 ANALYSOR	
Serial No:	A11124 / A106337	
Accessories:	FULL SET	
Quantity:	1	
Condition:	USED	
Work required:	WILL NOT CALIBRATE WOULDN'T GO BELOW 90	
Warranty:	Yes / No 16/11/23	PO Number: 130568
THIS WAS A WARRANTY REPLACEMENT		
Quotation required?	Yes / No	
Supplier action required:	Repair / Replace / Credit	