



COMPANY: Viamed Ltd

ADDRESS:
Returns Dept – SRS68905
Viamed Ltd
15 Station Road
Cross Hills
Keighley
BD20 7DT

Department of Medical Engineering & Physics
Kings College Hospital
Faraday Building
124-126 Denmark Hill
LONDON SE5 8RX

Tel: 020 3299 3538

DATE: 23/09/2024 TEL: +44 (0)1535 634542 Email: technical@viamed.co.uk

Dear Sir or Madam,

Please find enclosed the following equipment:

DESCRIPTION: Oxygen Monitor		
Model: MX300	Serial No: 340518	
ENGINEER: Radeet Kabir	WO. No: 325289	ID Code: 77-HM
PROBLEM: <ul style="list-style-type: none">- The Buttons do not function properly. It does not allow to set the High & Low Oxygen Limits.- The Buttons for the Battery & Alarm Tests do not work.- Whenever the Buttons are Locked & Unlocked it leads straight to Error 6.		
COMMENTS: <ul style="list-style-type: none">- Returns Reference: SRS68905- Decontamination Certificate attached inside.		

We would be grateful if you could:-

Provide an estimate for its repair: EMAIL TO - ENGINEERS EMAIL: radeet.kabir@nhs.net and kch-tr.memsrepairquotes@nhs.net	✓
Repair it against our official order	
Repair it under warranty	

ORDER No:		REQUISITION No:	
------------------	--	------------------------	--

Yours sincerely,

ALL EQUIPMENT TO BE RETURNED TO THE ADDRESS BELOW

Medical Engineering & Physics – MEMS
Faraday Building
124-126 Denmark Hill
LONDON
SE5 8RX
Tel: Direct Line 020 3299 3538