

Service Repair Sheet SRS68921

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Chantelle Louise Curry

Auto Trade Company

07564625494

07564625494

CID27563

autotrade01@outlook.com

09/Oct/2024

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 07/Oct/2024 Steve Hardaker

07/Oct/2024 Steve Hardaker

Returning oxygen sensor 0110122 (serial number 136587) under warranty.

Their customer is adamant that it is failing in the machine but has not provided specific details, other than it reads 12mV which should be fine. Trying to find more.

09/Oct/2024 Robert Connor

Received 1 x R22A s/n 136587, with threaded cap for membrane end.

Ready For quote

Repair Complete Signed

CC Connor 10-10-24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37128	Oxygen Sensor	0110122	136587	Yes

0140013 x 1 FOC
Old S/N, SRS, SRN

0110122 x 1 FOC
New S/N, SRS, SRN

UPS x 1 FOC