

DECLARATION OF CONFORMITY

Description: **DB200381 Oxygen Sensor**
Part number: **8010050**

Manufacturer: **Vandagraph Sensor Technologies Ltd.**
15 Station Road
Cross Hills
Keighley
West Yorkshire BD20 7DT
United Kingdom

Specification: **OEM Specification Datasheet – 8010050, v2.0 dated 6th May 2019**

Standards: **ISO 9001-2015**

JFD purchase order number: 12403914-2

Serial number/s	Manufacture – Year/Month
106894– 106910 (qty 17)	2024/09

I, the undersigned, hereby declare that the oxygen sensors listed above conform to the stated specifications.

Signature:

Name:


S Nixon

Position:

Director

Date:

20th September 2024



VERIFIED
CERTIFICATE

CHECKED BY **CW** DATE **02/10/24**
RECEIVER # **307280** OCR # **MULTIPLE**
WO/PO # **P012403914/1**
JFD PART NUMBER **DB200381**
TRACEABLE MARK **SERIALS 106894**
THRU 106910

Purchase order

12403914-2

< S_NT> S_NT

Date	Part number	Part number	Serial number	Output	Zero offset voltage	
	VST	JFD		mV	μV	
04/09/2024	8010050	DB200381	106894	7.8	28	2.78
04/09/2024	8010050	DB200381	106895	7.7	14	3.31
04/09/2024	8010050	DB200381	106896	7.8	34	3.67
04/09/2024	8010050	DB200381	106897	7.8	19	3.40
04/09/2024	8010050	DB200381	106898	7.8	16	3.69
04/09/2024	8010050	DB200381	106899	7.7	15	4.06
04/09/2024	8010050	DB200381	106900	7.9	11	3.09
04/09/2024	8010050	DB200381	106901	7.9	1	3.11
04/09/2024	8010050	DB200381	106902	7.9	33	3.01
04/09/2024	8010050	DB200381	106903	7.9	33	4.58
04/09/2024	8010050	DB200381	106904	7.9	16	4.10
04/09/2024	8010050	DB200381	106905	7.7	18	4.53
04/09/2024	8010050	DB200381	106906	7.8	13	3.78
04/09/2024	8010050	DB200381	106907	7.8	12	4.05
04/09/2024	8010050	DB200381	106908	7.7	17	5.09
04/09/2024	8010050	DB200381	106909	7.8	27	3.45
04/09/2024	8010050	DB200381	106910	7.9	23	4.41

Invoice number: RST150930
Shipment date: 20th September 2024
JFD reference: 12403914-2

**Magnetic signature analysis of production oxygen sensors
through JFD QA goods-in procedures**

17 x 8010050 Oxygen Sensor – JFD DB200381

Once the analysis has been carried out, please complete the following and email back to VST:

info@vandagraphst.com

Result (PASS/FAIL):

QTY 16 PASS 25mT

QTY 1 FAIL 75mT, AS HIGHLIGHTED

OVERLEAF

5.09mT

Name: CALLUM WALLS
Position: QUALITY ASSURANCE ENGINEER

Signed by:
On behalf of JFD



Date: 02 OCT 2024

Vandagraph Sensor Technologies Limited
15 Station Road, Cross Hills, Keighley, West Yorkshire BD20 7DT, United Kingdom
Tel +44 (0) 1535 634900 Fax +44 (0) 1535 635582 e-mail: info@vandagraphst.com
www.vandagraphst.com