



Kate Griffiths <viamed.kate.griffiths@gmail.com>

Re: Payment details

1 message

Sudath Weeraman <sudathweeraman17@gmail.com>
To: Kate Griffiths <kate.griffiths@viamed.co.uk>

16 October 2024 at 11:53

Good morning Kate,

The 1 Pc of MAX-250E Oxygen sensor is defective, S/N -JL55299066, the fault is varying the oxygen percentage

Please send return reference number for this under warranty,

Already informed to Skycargo express, for the collecting Flow sensor package,

Thank you,
Best regards,

Sudath
Prime Lanka Associates
Sri Lanka.

On Tue, Oct 15, 2024 at 4:26 PM Kate Griffiths <kate.griffiths@viamed.co.uk> wrote:

Hi Sudath,

Please see attached invoice.

The weights and dimensions are as follows:

31 x 24 x 24 cm

1.0 kg

Kind regards

Kate Griffiths

<http://www.viamed.co.uk>

Email kate.griffiths@viamed.co.uk

Tel: +44 (0)1535 634542

Fax: +44 (0)1535 635582

Twitter: twitter.com/ViamedLtd

Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.