

## Service Repair Sheet 68906

Contact Name	Matthew Jameson
Company/ Hospital Name	Sunderland Royal Hospital
Department	Clinical Engineering
Position	Senior Clinical Engineering Technician
Direct Phone	01915699933
General Phone	01915656256
Opera Account	00004992
Email	matthew.jameson@nhs.net
Order Number	RLNP400109116
Date Received	25/Sep/2024
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN37102	0310030	040022	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		

Tom Thumb TT480 040022 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

---

Repair	Ref	S/N	Equipment Type	
SRN37103	0310030	040092	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		

Tom Thumb TT480 040092 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

---

Repair	Ref	S/N	Equipment Type	
SRN37104	0310030	VMA0077	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		

Tom Thumb TT480 VMA0077 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

---

Repair	Ref	S/N	Equipment Type	
SRN37105	0310030	VMA0081	Tom Thumb	Complete - Repaired Time :0 Hour(s)
Parts Replaced				
	Qty: 1	0380000		

Tom Thumb TT480 VMA0081 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

---