

Service Repair Sheet 68906

Contact Name	Matthew Jameson
Company/ Hospital Name	Sunderland Royal Hospital
Department	Clinical Engineering
Position	Senior Clinical Engineering Technician
Direct Phone	01915699933
General Phone	01915656256
Opera Account	00004992
Email	matthew.jameson@nhs.net
Order Number	RLNP400109116
Date Received	25/Sep/2024
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37106	0310030	VMA0083	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 VMA0083 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN37107	0310030	VMA0087	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 VMA0087 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN37108	0310030	VMA106	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0330236
Qty: 0.25 0340000

Tom Thumb TT480 VMA106 has been stripped and cleaned then reassembled with new O rings and a new gauge face. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	
SRN37109	0310030	VMA104	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 VMA104 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.
