

Box 1001
requested by
email 11/10/24

Service Repair Sheet SRS68916

Contact Name Ashley Davies
Company/ Hospital Name Thames Medical Ltd
Department
Position
Direct Phone 01903522911
General Phone 01903522911
Opera Account 00011982
Email ashley@thamesmedical.com
Order Number
Date Received 07/Oct/2024
Booked in By Robert Connor
Main Company Viamed
Type Return Quote

VIAMEDclean	
Goods In Only	
Decontamination certificate provided by customer	<input checked="" type="checkbox"/>
Cleaned by Viamed, if no declaration certificate from customer <input type="checkbox"/>	
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes Power upgrade required for VM-2500M, s/n 81030155. They also require power supply and UK plug adaptor. They are aware 5-6 week lead time

03/Oct/2024 Kate Griffiths

Power upgrade required for VM-2500M, s/n 81030155. They also require power supply and UK plug adaptor. They are aware 5-6 week lead time

07/Oct/2024 Robert Connor

Received 1 x VM-2500-M-Vet s/n 81030155, pin broken in power connector, with li-poly battery and rev-2 power supply and UK adapter, pin in power supply.

Ready For quote

CGreen 11.10.24

Repair Complete Signed

CGreen 25.10.24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37124	Capnograph	4410530	81030155	N

4440000 x 0.50hr @ £90 = ~~£450~~ £45
S/N, SRS, SRN

4420598 x 1 @ £34.35
SRS, SRN

4430607 x 1 @ £5.10
SRS, SRN

4400080 x 1 @ £42.76
SRS, SRN

UPS x 1 @ £12.