

Service Repair Sheet 68929

Contact Name	Ross Bowden
Company/ Hospital Name	Cardiff and Vale UHB
Department	Clinical Engineering
Direct Phone	02921845678
General Phone	02920903794
Opera Account	00000092
Email	Ross.Bowden@wales.nhs.uk
Order Number	726670150
Date Received	15/Oct/2024
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37132	1410000	PR02450A13	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1 1430309

Qty: 1 1480000

V1000 has had transducer interface cushion replaced and calibration has been checked.