

# Service Repair Sheet SRS68904

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Alex Ferguson

Churchill Hospital

Clinical Engineering

Tech

01865272163

01865 741841

00004030

Alex.Ferguson@ouh.nhs.uk

24/Sep/2024

Robert Connor

Viamed

Quote

## VIAMEDclean

### Goods In Only

Decontamination  
certificate provided  
by customer



Cleaned by Viamed,  
if no declaration certificate  
from customer

Signed: \_\_\_\_\_  
Date: \_\_\_\_\_



### Goods Out Only

Cleaned by Viamed before  
returning to customer

Signed: \_\_\_\_\_  
Date: \_\_\_\_\_

Notes 20/Sep/2024 Aqib Majeed

20/Sep/2024 Aqib Majeed

customer returning 2 x Microstim DB3 nerve stimulators Sn: M0005195 & Sn: M0005194 for repair, Both devices have no signal output despite trying new patient leads and batteries

24/Sep/2024 Robert Connor

Received 2 x Microstim DB3 s/n M0005194, M0005195, missing one battery door.

Ready For quote

CGNeen 2-10-24

Repair Complete Signed

CGNeen 7-11-24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37099	Microstim DB3	2510000	M0005194	22
SRN37100	Microstim DB3	2510000	M0005195	22

2540010 x 2 @ £60 each

SIN's, SRS, SEN's

UPS x 1 @ £12.