

Service Repair Sheet SRS68931

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Sudath Weeraman

Prime Lanka Associate

9410718732285

00006714

sudathweeraman17@gmail.com

06/Nov/2024

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 16/Oct/2024 Kate Griffiths

16/Oct/2024 Kate Griffiths

1 x MAX-250E 0110429 to be returned under warranty - ``the fault is varying the oxygen percentage``. S/N: JL55299066

06/Nov/2024 Robert Connor

Received 1 x MAX-250E s/n JL55299066, with flow diverter.

Ready For quote

Repair Complete Signed

C. Green 11-11-24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37184	Oxygen Sensor	0110429	JL55299066	Yes

0146013 x 1 FOC
old S/N, SRS, SRN

0110429 x 1 FOC
New S/N, SRS, SRN

Exw..
Ship with next Order