



Service Repair Sheet SRS68940

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Mark Evans
Royal United Hospitals Bath
Medical Equipment Management
Senior Clinical Technologist
01225824140
01225428331
00000350
mark.evans7@nhs.net

12/Nov/2024

Robert Connor

Viamed

Warranty

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____
Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 28/Oct/2024 Kate Griffiths

28/Oct/2024 Kate Griffiths

would like to return a faulty MAX 550e, s/n HH82199042 - it keeps changing from 1 to 13. Under warranty

12/Nov/2024 Robert Connor

Received 1 x MAX-550E s/n HH82199042, with flow diverter. Output in air drops when changing orientation.

Ready For quote

Repair Complete Signed

CGNORR 14.11.24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37187	Oxygen Sensor	0110452	HH82199042	Yes

0140013 x 1 FOC

old S/N, SRS, SRN

0110452 x 1 FOC

new S/N, SRS, SRN

UPS x 1 ~~FOC~~ FOC