

Dispatch note - External service

Our reference	288269	Supplier	VIAMED LTD
Order date	28/10/2024	Address	15 STATION ROAD, CROSS HILLS BD20 7DT KEIGHLEY, W YORKSHIRE
Our reference	WO: 288269	Your reference	SRS68940
Contact person	MARK EVANS	Phone	01535 634542
Phone	01225 824154	Fax	01535 635582
Fax		Package no	1
Mobile phone		Type of service	WARRANTY WORK
Email	mark.evans7@nhs.net		

MEMS no.	49105	Serial no	FM72099005
Device type	OXYGEN MONITOR		
Brand	MAXTEC		
Model	MAX O2 ME		
Owner	0707 NICU ; RUH		
Location	0707 NICU ; RUH		

*The serial number is
The monitor
the sensor it
came out of.
Sensor only received.
checked with Mark Evans
over phone today
not expecting eg 18.11.24
the main unit.*

Message
Hi

Please repair or replace this oxygen cell. The percentage of O2 measured in air can drop to 13% or 0% depending whether the sensor is pointing up or down.

Many thanks

Mark Evans

Return date

28/10/24

Signature

Mark Evans

Before commencing an in-depth repair, could you please advise us by telephone of the likely cost if over £
0.00 by asking for the contact person above. This letter serves as a certificate that the device has been
decontaminated to standard local protocols. PLEASE RETURN ITEM TO DELIVERY ADDRESS BELOW

The goods are to be returned to the address below

Invoice address Finance Dept, Malvern House -
Building 6
Royal United Hospital
Combe Park
BATH
BA1 3NG

Delivery address MEMS - Department A8
ROYAL UNITED HOSPITAL
COMBE PARK
BATH
BA1 3NG

**This item has been
decontaminated**

Signature

Signed
Date

Mark Evans
28/10/24