

Service Repair Sheet 68956

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

David Watson

Darlington Memorial Hospital

Clinical Engineering Department

Clinical Engineering Technician

01325743048

01325380100

00001350

david.watson28@nhs.net

CDD0006180123

13/Nov/2024

Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37350	0310030	VMBC19	Tom Thumb	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1 0380000

Tom Thumb TT480 VMBC19 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.