

Service Repair Sheet SRS68964

Contact Name

Tim Bray

Company/ Hospital Name

Medisafe Ltd

Department

Position

Direct Phone

01962762098

General Phone

01962885777

Opera Account

00011608

Email

tim.bray@medisafe.org.uk

Order Number

Date Received

20/Nov/2024

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer



Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 19/Nov/2024 Kate Griffiths

19/Nov/2024 Kate Griffiths

Tim rang. His V1000, s/n: PR0490A19A is not switching on. They will send it in. Asked for cert of decontamination and no batteries

20/Nov/2024 Robert Connor

Received 1 x V1000 s/n PR0490A19A, with 4 x AA battery.

Ready For quote

CC, Neen 21.11.24

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37464	Foetal Simulator	1410000	PR0490A19A	N

1480000 x 1 @ £60

S/N, SRS, SRN.

1430309 x 1 £0

SRS, SRN

UPS x 1 £12.