

eg 20.8.24

Service Repair Sheet SRS68855

Contact Name	Atif Hussain
Company/ Hospital Name	Luton & Dunstable Hospital
Department	Clinical Engineering Department
Position	Theatre Equipment Manager
Direct Phone	01582491166 ext 2480
General Phone	01582497042
Opera Account	00003490
Email	atif.hussain@ldh.nhs.uk
Order Number	SRS68855
Date Received	29/Jul/2024
Booked in By	Robert Connor
Main Company	Viamed
Type Return	Quote

<b>VIAMEDclean</b>	
<b>Goods In Only</b>	
Decontamination certificate provided by customer	<input type="checkbox"/>
Cleaned by Viamed, if no declaration certificate from customer	<input type="checkbox"/>
Signed: _____	
Date: _____	
<b>Goods Out Only</b>	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 23/Jul/2024 Aqib Majeed  
 23/Jul/2024 Aqib Majeed  
 customer has a Capnograph Device which displays sensor fault, SN number is A1711160006, no warranty on device.  
 29/Jul/2024 Robert Connor  
 Received 1 x VM-2500-S s/n A1711160006, with silicone cover.

Ready For quote CGreen 13-9-24

Repair Complete Signed CGreen 27-11-24

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN36564	Capnograph	4410520	A1711160006	N

4440000 x 1 1/2 hr @ £90 per hour

S/N, SRS, SRN

4480520 x 1 @ £150

SRS, SRN

4420504 x 1 @ £1,049.75

SRS, SRN

4430201 x 1 @ £15

SRS, SRN

UPS x 1 £12.