

Service Repair Sheet 68946

Contact Name	Sreelakshmy Mookola Rejendra Prasad
Company/ Hospital Name	Whipps Cross University Hospital
Department	Clinical Engineering Department
Position	Clinical Technologist/ Quality Rep.
Direct Phone	02085356927
General Phone	02085356927
Opera Account	00003450
Email	s.prasad2@nhs.net
Order Number	41084950
Date Received	12/Nov/2024
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN37188	1410000	PR01202A17	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.
