

Viamed Limited · 15 Station Road · Cross Hills Keighley · West Yorkshire · BD20 7DT · United Kingdom Tel: +44 (0)1535 634 542 · Fax: +44 (0)1535 635 582 Email: info@viamed.co.uk Website: www.viamed.co.uk

Service Repair Sheet 68968

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone General Phone

Opera Account

Email

Order Number Date Received Booked in By

Arlene Gallagher

Forth Valley Royal Hospital

Medical Physics Dept

Ms Gallagher 01324 566953

01324 566000 00004880

arlene.gallagher@nhs.scot

VMF14834763 27/Nov/2024 Robert Connor

Repair

Ref

S/N

SRN37470 1410000 PR01985A11

Equipment Type

Foetal Simulator

Complete - Repaired Time: 0 Hour(s)

Parts Replaced

Qty: 1 1430309 Qty: 1 1480000

V1000 has had transducer interface cushion replaced and calibration has been checked.