

Service Repair Sheet 68961

Contact Name	Ros Rochester
Company/ Hospital Name	Worcester Royal / Siemens Healthcare
General Phone	01905733176
Opera Account	00005446
Email	ros.rochester@siemens-healthineers.com
Order Number	4515581549
Date Received	27/Nov/2024
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN37467	0111263	GC48099075	Oxygen Monitor	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	R228P15
Qty: 0.5	0140120

The faulty keypad has been replaced and the calibration of the monitor GC48099075 has been checked.

---

Repair	Ref	S/N	Equipment Type	
SRN37468	0131700	Tracking22	Misc. Cable	No Fault Found Time :0 Hour(s)

---

Repair	Ref	S/N	Equipment Type	
SRN37469	0110452	JL10199094	Oxygen Sensor	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	0110452
--------	---------

Replaced with S/N KH85099003

---