

Equipment for Repair

Date: 5/12/24
To: VIANCO

Repair Reference - 70937
(Rev 61433)

Vet Direct
Units 9 & 10, Gateway West
Newburn Riverside
Newcastle upon Tyne
NE15 8NZ

Tel: 0191 2642705
Email: repairs@vet-direct.com

The enclosed product has been sent to you for inspection. Please provide estimate to repair - please send estimate to repairs@vet-direct.com

Product: OXIGANT 5 ANALYSER
Serial No: A110805 / A106106
Accessories: ADAPTORS
Quantity: 1
Condition: USED
Work required: UNABLE TO CALIBRATE
(PREVIOUSLY SENT BACK AND HAD
BUTTON PANEL REPAIRED)
Warranty: Yes / No ?? 2/5/23 PO Number:
Quotation required? ☒ Yes / ☐ No IF NOT UNDER WARRANTY
Supplier action required: ☒ Repair / ☐ Replace / ☐ Credit