

Service Repair Sheet SRS68975

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Louis Fretwell

St James University Hospital

Medical Engineering Dept

Medical Engineering Apprentice

01132433144

00002570

louis.fretwell@nhs.net

11/Dec/2024

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer



Cleaned by Viamed,
if no declaration certificate
from customer



Signed: _____

Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 05/Dec/2024 Steve Hardaker

05/Dec/2024 Steve Hardaker

Returning a Microstim DB3 Serial Number M0004338 that is no longer working. Has tried replacing the leads without success. Please process as a fixed cost repair.

11/Dec/2024 Robert Connor

Received 1 x DB3 s/n M0004338, with patient lead, intensity control knob detached.

Ready For quote

CGREEN 12.12.24

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37531	Microstim DB3	2510000	M0004338	2

2540010 x 1 £60

SIN, SRS, SRN

UPS x 1 £12.