

Service Repair Sheet 68963

Contact Name	Matthew Jameson
Company/ Hospital Name	Sunderland Royal Hospital
Department	Clinical Engineering
Position	Senior Clinical Engineering Technician
Direct Phone	01915699933
General Phone	01915656256
Opera Account	00004992
Email	matthew.jameson@nhs.net
Order Number	RLNP400113339
Date Received	19/Nov/2024
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN37452	0310030	040020	Tom Thumb	Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 040020 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN37453	0310030	040023	Tom Thumb	Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 040023 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN37454	0310030	040024	Tom Thumb	Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 040024 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.

Repair	Ref	S/N	Equipment Type	Complete - Repaired
SRN37455	0310030	040352	Tom Thumb	Time :0 Hour(s)

Parts Replaced
Qty: 1 0380000

Tom Thumb TT480 040352 has been stripped and cleaned then reassembled with new O rings. The two valves have been checked and reset before QA.
