

Service Repair Sheet SRS68972

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Ros Rochester
Worcester Royal / Siemens Health

01905733176
01905733176
00005446
ros.rochester@siemens-healthine

09/Dec/2024
Robert Connor
Viamed
Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer

Signed: _____
Date: _____

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____
Date: _____

Notes 03/Dec/2024 Kate Griffiths

03/Dec/2024 Kate Griffiths

Customer would like to send in 4 z MAXO2 ME 0111263 - s/n: GC48099076, GC48099078, GC48099080
GC48099079. Error codes 1 and 3, calib error. They tried to recalibrate with 20.9% and 100% O2 but still showing errors.
Tried new sensors, still same errors. Asked for cert of decont and batteries removed

09/Dec/2024 Robert Connor

4 x MaxO2ME s/n GC48099076, with coiled cable no s/n MAX550E s/n GC25399088, with flow diverter and blue tee.
s/nGC48099080, with coiled cable s/n 1748707, MAX550E s/n GC25399084, with flow diverter and blue tee. s/n
GC48099078, with coiled cable s/n 1748702, MAX550E s/n GC25399089, and flow diverter. s/n GC48099079, with coiled
cable s/n unreadable, MAX550E s/n GC25399082, with flow diverter and blue tee.

Ready For quote

C. Green 6.1.25.

Repair Complete Signed _____

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37476	Oxygen Monitor -	0111263	GC48099076	2
SRN37477	Oxygen Monitor -	0111263	GC48099078	2
SRN37478	Oxygen Monitor -	0111263	GC48099079	2
SRN37479	Oxygen Monitor -	0111263	GC48099080	2
SRN37480	Oxygen Sensor	0110452	GC25399082	2
SRN37481	Oxygen Sensor	0110452	GC25399084	2
SRN37482	Oxygen Sensor	0110452	GC25399088	2
SRN37483	Oxygen Sensor	0110452	GC25399089	2
SRN37484	Misc. Cable	0131700	Tracking12	2
SRN37485	Oxygen Monitor Cable	0131700	Tracking17	2
SRN37486	Oxygen Monitor Cable	0131700		2
SRN37487	Oxygen Monitor Cable	0131700		2

} Part of
units

0140000
~~0040100~~ x 4 @ £0.00 NO fault found
S/N'S SRS, SRSN'S → SRN 37476-37479

0110452 x 4 @
SRS, SRN E106-60
each.

0140015 x 4
→ SRN 37480-SRN37483

0140025 x 4 FCC
NO fault. → SRN 37484
SRS-SRN → SRN 37487