

QA
0310200 Low-Flow Blender air & oxygen.
Barcode ID: 2651432
Serial Number BX1854643

Carried out by: Philip Crossley
07/Jan/2025 3:17 pm

Performance evaluation: Pass

Signed: _____

Name: _____

Position: _____

Date: _____



Viamed Limited · 15 Station Road · Cross Hills
Keighley · West Yorkshire · BD20 7DT · United Kingdom
Tel: +44 (0)1535 634542/636757 Fax: +44 (0)1535 635582
Email: info@viamed.co.uk Website: www.viamed.co.uk

QA 0310200 Serial Number BX1854643 Barcode ID: 2651432
Carried out by: Philip Crossley 07/Jan/2025 3:17 pm

[1]2303 Asset Number
RES030A

[2]2206 21% Output between 20.8 - 23
21.2

[3]2205 30% Output between 27 - 33
27.9

[4]2204 60% Output between 57 - 63
61.3

[5]2203 90% Output between 87 - 93
91.6

[6]2202 100% Output between 98-100
100.0

[7]2201 Check Alarm for no Oxygen
Yes

[8]2199 Check Alarm for no Air
Yes

[9]2198 Check O2/Air Balance Alarms before output change
Yes

[10]2197 Check Duck Bill Valves
Yes

[11]2196 Check AUX
Yes

[12]2195 Check for Leaks
Yes