

Service Repair Sheet SRS68893

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Anna Carlin

Swevet - AB

Purchase Assistant

46771215500

46771215500

00006308

anna.carlin@swevet.se

19/Nov/2024

Robert Connor

Viamed

Quote

VIAMEDclean

Goods In Only

Decontamination
certificate provided
by customer ☒

Cleaned by Viamed,
if no declaration certificate
from customer ☒

Signed: RS

Date: 19/11/24

Goods Out Only

Cleaned by Viamed before
returning to customer

Signed: _____

Date: _____

Notes 06/Sep/2024 Ryan Swaine

06/Sep/2024 Ryan Swaine

VM-2160 VET device with a Y sensor

sn. T2201110025

Under warranty.

Quote from customer:

The customer could not get it to measure properly. The dog sat still but worked for a few seconds and then didn't. They have tried tongue, paw and ear. Works on fingertip but not on animals. They have tried several animals.

19/Nov/2024 Robert Connor

Received 2 x VM-2160-Vet s/n T2201110025, with carry case, lanyard, silicone cover, USB cable, Y probe clip, 3 x AA battery, Y probe s/n VYV210039, s/n T2201110022, with carry case, silicone cover, 2 x Y probe clip, Y probe s/n VYV210035.

Ready For quote

Repair Complete Signed

C Green 14.1.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37460	Pulse Oximeter	0012176	T2201110022	
SRN37461	Pulse Oximeter	0012176	T2201110025	Yes
SRN37462	Pulse Oximetry Y Sensor	0015002	VYV210035	Yes
SRN37463	Pulse Oximetry Y Sensor	0015002	VYV210039	Yes

0140050 x 2 FOC
S/N's SRS, SRN'S

0040001 x 2 FOC
S/N's SRS, SRN'S

005 x 1 FOC.