

Service Repair Sheet SRS68972

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Ros Rochester
Worcester Royal / Siemens Health

01905733176
01905733176
00005446
ros.rochester@siemens-healthine

09/Dec/2024
Robert Connor
Viamed
Quote

VIAMEDclean	
Goods In Only	☒
Decontamination certificate provided by customer	
Cleaned by Viamed, if no declaration certificate from customer	☒
Signed: _____	
Date: _____	
Goods Out Only	
Cleaned by Viamed before returning to customer	
Signed: _____	
Date: _____	

Notes 03/Dec/2024 Kate Griffiths

03/Dec/2024 Kate Griffiths

Customer would like to send in 4 z MAXO2 ME 0111263 - s/n: GC48099076, GC48099078, GC48099080
GC48099079. Error codes 1 and 3, calib error. They tried to recalibrate with 20.9% and 100% O2 but still showing errors.
Tried new sensors, still same errors. Asked for cert of decont and batteries removed

09/Dec/2024 Robert Connor

4 x MaxO2ME s/n GC48099076, with coiled cable no s/n MAX550E s/n GC25399088, with flow diverter and blue tee.
s/nGC48099080, with coiled cable s/n 1748707, MAX550E s/n GC25399084, with flow diverter and blue tee. s/n
GC48099078, with coiled cable s/n 1748702, MAX550E s/n GC25399089, and flow diverter. s/n GC48099079, with coiled
cable s/n unreadable, MAX550E s/n GC25399082, with flow diverter and blue tee.

Ready For quote

Cgreen 6.1.25.

Repair Complete Signed

Cgreen 15.1.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37476	Oxygen Monitor	0111263	GC48099076	2
SRN37477	Oxygen Monitor	0111263	GC48099078	2
SRN37478	Oxygen Monitor	0111263	GC48099079	2
SRN37479	Oxygen Monitor	0111263	GC48099080	2
SRN37480	Oxygen Sensor	0110452	GC25399082	2
SRN37481	Oxygen Sensor	0110452	GC25399084	2
SRN37482	Oxygen Sensor	0110452	GC25399088	2
SRN37483	Oxygen Sensor	0110452	GC25399089	2
SRN37484	Misc. Cable	0131700	Tracking12	2
SRN37485	Oxygen Monitor Cable	0131700	Tracking17	2
SRN37486	Oxygen Monitor Cable	0131700		2
SRN37487	Oxygen Monitor Cable	0131700		2

} Part of units

0140000
0040100 x 4 @ £0.00 NO fault found
S/N'S SRS, SEN'S -> SEN 37476-37479
0140015 x 4 -> SEN 37480-SEN37483

0110452 x 4 @ 103.90
NEW S/N SRS, SRN £106.60 each.
CG 15.1.25 -> Addl.
0140025 x 4 FOC
NO fault -> SEN 37484
SRS-SEN