

Service Repair Sheet SRS68947

Contact Name Avram Bratspiess
Company/ Hospital Name I.M.T.C Ltd
Department
Position
Direct Phone
General Phone 97286636637
Opera Account 00007088
Email avram@imtc.co.il
Order Number
Date Received 10/Dec/2024
Booked in By Robert Connor
Main Company Viamed
Type Return Warranty

Notes 04/Nov/2024 Ryan Swaine

04/Nov/2024 Ryan Swaine

Unknown qty of R-22MED sensors which are failing with reportedly low or no output. The end user is blaming the age of the sensors, so sensors are to be exchanged with R-22MED sensors that are within 4 months of manufacture date.

10/Dec/2024 Robert Connor

Received 37 x R22MED, customer delivery note states 41, however some may have been lost in transit due to careless packing by customer.

Ready For quote

Repair Complete Signed

CGNOR 17-1-25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37494	Oxygen Sensor	0110022	956250	Yes .
SRN37495	Oxygen Sensor	0110022	956251	Yes .
SRN37496	Oxygen Sensor	0110022	956256	Yes .
SRN37497	Oxygen Sensor	0110022	956258	Yes .
SRN37498	Oxygen Sensor	0110022	956260	Yes .
SRN37499	Oxygen Sensor	0110022	956261	Yes .
SRN37500	Oxygen Sensor	0110022	956262	Yes .
SRN37501	Oxygen Sensor	0110022	624794	X
SRN37502	Oxygen Sensor	0110022	786491	X
SRN37503	Oxygen Sensor	0110022	875102	X
SRN37504	Oxygen Sensor	0110022	935869	Yes .
SRN37505	Oxygen Sensor	0110022	935874	Yes .
SRN37506	Oxygen Sensor	0110022	935877	Yes .
SRN37507	Oxygen Sensor	0110022	935883	Yes .
SRN37508	Oxygen Sensor	0110022	935884	Yes .
SRN37509	Oxygen Sensor	0110022	935947	Yes .
SRN37510	Oxygen Sensor	0110022	935948	Yes .
SRN37511	Oxygen Sensor	0110022	935949	Yes .
SRN37512	Oxygen Sensor	0110022	935961	Yes .
SRN37513	Oxygen Sensor	0110022	935981	Yes .
SRN37514	Oxygen Sensor	0110022	935982	Yes .

PTO.