

## Service Repair Sheet 69003

Contact Name Vincent Ochieng  
Company/ Hospital Name University College Hospital  
Direct Phone 02034474984  
General Phone 0845 1555000/02034 567890  
Opera Account 00003420  
Email v.ochieng@nhs.net  
Date Received 15/Jan/2025  
Booked in By Robert Connor

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN37580	1410000	PR03485A13	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

## Parts Replaced

Qty: 1	1480000
Qty: 1	1430309

V1000 has had transducer interface cushion replaced and calibration has been checked.