



Kate Griffiths <viamed.kate.griffiths@gmail.com>

Re: O2 cell & flow sensor for Drager and Hamilton

1 message

Kate Griffiths <kate.griffiths@viamed.co.uk>

27 January 2025 at 11:59

To: Ojie <sales1@whemd.com>

Bcc: Main Account <viamedinbox@gmail.com>

Good morning Ojie,

Thank you for your email.

1. Part number 4310001 has been **superseded by part number 4310007** - Flowsensor A - single use flow sensor, box of 5. We have transitioned over to the Bluepoint branded versions. **These are the same product as the Viamed branded.**
2. Part number 0110023 - oxygen sensor R-23V is our sensor for the Babylog 800 Model: Evita
3. Part number 4310009 - Flowsensor H - Single use flow sensor, box of 6 is for Hamilton Medical H7400. **This is a box of 6.**
4. Hamilton Oxygen cell for C2, Raphael, C5, Gallelio, etc. - this may be a no, but we can supply the R-49V/0110049 for Raphael/G5 - please see attached specification.
5. Flow sensor for SLE vents, N5402-REV2. **We are unable to supply this item.**
6. Flow sensor Cable for SLE vents, N6656. **We are unable to supply this item.**
7. Part number 0110590 - Oxygen sensor OOM111-M for Maquet/Servo-i

After you have looked at my email/attachments, please could you let me know which items you require on an official quotation?

Many thanks.

Kind regards

Kate Griffiths

<http://www.viamed.co.uk>

Email kate.griffiths@viamed.co.uk

Tel: +44 (0)1535 634542

Fax: +44 (0)1535 635582

Twitter: twitter.com/ViamedLtd

Facebook Page: Search for Viamed Ltd

Company registered in England, No. 01291765.

Viamed is now enacting hybrid working to allow sales and admin staff to work remotely and in the office. Telephone calls to the main office will be answered, but please continue to use email where possible.

E-mail Disclaimer

GDPR / Privacy Policy available at www.viamed.co.uk/policy

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

On Sat, 25 Jan 2025 at 07:47, Ojie <sales1@whemd.com> wrote:

Dear Kate,

Pls. kindly quote us the following for Drager & Hamilton, as follows;

- | | |
|--|-------------|
| 1. Drager flow sensor, #8403735 | 5bx |
| Viamed P/N: 4310001 | |
| 2. Drager O2 cell for the following model | 10 ea model |
| Model : VN 500 | |
| Model : Babylog 8000 | |
| Model : Evita | |
| 3. Hamilton Flow sensor ,disposable w/tubing, # H7400 | 5 bx |
| 4. Hamilton Oxygen cell for following Model; | 5 ea model |
| C2, Raphael, C5,Gallelio, etc..(pls. advise if uses the same O2 sensor) | |
| 5. Flow sensor for SLE vents, N5402-REV2 | 30 ea |
| 6. Flow sensor Cable for SLE vents, N6656 | 20 ea |
| 7. Oxygen Cell for Maquet vents, Servoi | 10 ea |

Looking forward to hear from you soon.

Thanks & Best Regards,



Ojie Miranda

Manager- Respiratory Department

World Health Excellence Medical Company

0138941919 Mob: 0507894157

sales1@whemd.com

<https://www.whemd.com>

7484, King Khalid St, 34427 Al Khobar, Kingdom of Saudi Arabia

4 attachments



SpiroTrue Bluepoint Flowsensor Leaflet ID119548.pdf

310K



SpiroTrue Bluepoint Flowsensor Leaflet ID119550.pdf

846K



Bluepoint Medical - Leaflet - SpiroTrue H ID124876.pdf

61K



0110049 Specification Datasheet ID17622.pdf

200K