

Contact Name  
Company/ Hospital Name  
Department  
Position  
Direct Phone  
General Phone  
Opera Account  
Email  
Order Number  
Date Received  
Booked in By  
Main Company  
Type Return

Louis Fretwell  
St James University Hospital  
Medical Engineering Dept  
Medical Engineering Apprentice

01132433144  
00002570  
louis.fretwell@nhs.net

11/Dec/2024  
Robert Connor  
Viamed  
Quote

**VIAMEDclean****Goods In Only**

Decontamination  
certificate provided  
by customer



Cleaned by Viamed,  
if no declaration certificate  
from customer



Signed: \_\_\_\_\_

Date: \_\_\_\_\_

**Goods Out Only**

Cleaned by Viamed before  
returning to customer

Signed: \_\_\_\_\_

Date: \_\_\_\_\_

Notes 05/Dec/2024 Steve Hardaker

05/Dec/2024 Steve Hardaker

Returning a Microstim DB3 Serial Number M0004338 that is no longer working. Has tried replacing the leads without success. Please process as a fixed cost repair.

11/Dec/2024 Robert Connor

Received 1 x DB3 s/n M0004338, with patient lead, intensity control knob detached.

Ready For quote

CGreen 12.12.24

Repair Complete Signed

CGreen 24.1.25

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN37531	Microstim DB3	2510000	M0004338	2

2540010 x1 £60  
SIN, SRS, SRN

UPS x1 £12.