Service Repair Sheet SRS69012

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received Booked in By

Main Company Type Return

Richard Wood

University Hospital Monklands

01698752074

00000070

Richard.Woods@lanarkshire.scot

29/Jan/2025

Robert Connor

Viamed

Quote

VIAMEDclean Goods In Only

Decontamination certificate provided

by customer Cleaned by Viamed, if no declaration certificate from customer Signed: Date: -

Goods Out Only

Cleaned by Viamed before returning to customer

Signed:

Notes customer wants to return faulty V1000 device, said outputs are low 27/Jan/2025 Agib Majeed

Serial number for device is PR02894A14

29/Jan/2025 Robert Connor

Received 1 x V1000 s/n PR02894A14, with 4 x AA battery and hard plastic carry case.

SRN37593 Foetal Simulator

Ready For quote

cGreen 30-1.25

PR02894A14

Repair Complete Signed

| Equipment | 01.1.0.4 | Serial Number | 10/0 | |
|-----------|----------|---------------|------|--|

1410000

1480000 cg 30-1-25 SIN, SES, SEN

SRN

1430309 × 1 FCC 5RS, SEN

68×10 £12.